

# MAGLUMI CSA ( CLIA )

REF

130203002M



100



**Shenzhen New Industries  
Biomedical Engineering Co., Ltd**  
4F, Wearnes Tech Bldg,  
Science & Industry Park,  
Nanshan, Shenzhen, 518057 CHINA  
Tel. + 86 755 26508518  
Fax. + 86 755 26654850

EC REP

**Lotus Global Co., Ltd**  
15 Alexandra Road  
London  
NW8 0DP  
UK  
Tel. + 44-20-75868010  
Fax. + 44-20-79006187



## FOR PROFESSIONAL USE ONLY

Store at 2...8 °C



COMPLETELY READ THE INSTRUCTIONS BEFORE  
PROCEEDING

en

## SYMBOLS EXPLANATIONS

EC REP

Authorized Representative in Europe



Manufacturer



Attention. See Instructions For Use

CONT

Contents of kit

IVD

In vitro diagnostic medical device  
(In vitro diagnostic use)

LOT


Lot number

REF

Catalogue Code



Expiry date (Use by...)



Temperature limitation  
(store at 2...8 °C)



Number of tests



Keep away from sunlight



Keep upright

## INTENDED USE

The kit has been designed for the quantitative determination of Cyclosporine A(CSA) in human serum.

The method can be used for samples over the range of 0-2000 ng/ml.

The test has to be performed on the MAGLUMI fully auto analyzer (Including MAGLUMI 1000, MAGLUMI 2000, MAGLUMI 2000 Plus).

## SUMMARY AND EXPLANATION OF THE TEST

Cyclosporine A is a cyclic undecapeptide of fungal origin and a potent immunosuppressant.

It is used as a primary agent during immunosuppressive therapy of solid organ transplants. Immunosuppression is thought to be the result of impairment of T-cell receptor transcription of the IL-2 gene. Cyclosporine A therapy has greatly improved organ transplant survival of skin, heart, kidney, pancreas, bone marrow, lung, small intestine, and liver transplants.

Cyclosporine A may be administered by IV or orally. Absorption from the gastrointestinal is variable, unpredictable, and incomplete. Bioavailability increases during treatment so oral doses must be gradually reduced in order to maintain a constant cyclosporine A concentration in the blood. Assessment of cyclosporine A concentrations in blood aids in adjusting patients' dosage and avoids cyclosporine A underdosage inefficacy or overdosage toxicity. Cyclosporine A is eliminated almost completely by hepatic metabolism; cytochrome P-450 enzymes being responsible for the biotransformation of cyclosporine A and its metabolites. More than thirty metabolites have been identified. Preliminary data indicate cyclosporine metabolites are less immunosuppressive and less toxic than their parent compound. Many drugs affect cyclosporine A blood concentrations. These drugs alter cyclosporine A blood concentration by inducing drug metabolism, interfering with drug metabolism, or affecting drug absorption. Such interactions between cyclosporine A and danazol, diltiazem, erythromycin, fluconazole, itraconazole, ketoconazole, metoclopramide, nicardipine, verapamil, carbamazepine, phenobarbital, phenytoin, rifampicin, and cotrimoxazole are well documented.

The use of cyclosporine A is associated with serious toxic side effects, primarily nephrotoxicity and hepatotoxicity. Other adverse effects include diarrhea, gum hyperplasia, nausea, vomiting, hirsutism, tremor, and hypertension.

## PRINCIPLE OF THE TEST

Competitive immunoluminometric assay;

Use purified Cyclosporine A antigen to label ABEL, and anti-Cyclosporine A polyclonal antibody to label microbeads.

Sample, Calibrator or Control, ABEL Label, and magnetic microbeads coated with anti-Cyclosporine A polyclonal antibody are mixed thoroughly and incubated at 37°C, forming antibody-antigen complexes; after sediment in a magnetic field, decant the supernatant, then cycle washing for 1 time.

Subsequently, the starter reagents are injected and a flash chemiluminescent reaction is initiated. The light signal is measured by a photomultiplier as RLU within 3 seconds and is proportional to the concentration of Cyclosporine A present in controls or samples.

CONT

## KIT COMPONENTS

Material Supplies

Reagent Integral for 100 determinations	
Nano magnetic microbeads : TRIS buffer, 1.2%(W/V), 0.2%NaN <sub>3</sub> , coated with anti-Cyclosporine A polyclonal antibody.	2.5ml
Calibrator low: bovine serum, 0.2%NaN <sub>3</sub> .	3.0ml
Calibrator high: bovine serum, 0.2%NaN <sub>3</sub> .	3.0ml
ABEI Label : purified Cyclosporine A antigen labeled ABEI, containing BSA, 0.2%NaN <sub>3</sub> .	10.5ml
0.30M EDTA	5.5ml
0.83%NH <sub>4</sub> Cl	25ml
All reagents are provided ready-to-use.	

Please check the chapter: **SPECIMEN COLLECTION AND PREPARATION** carefully for the preparation of sample.

Reagent Vials in kit box	
Internal Quality Control: containing BSA, 0.2%NaN <sub>3</sub> , (target value refer to Quality Control Information date sheet)	2.0ml

#### Accessories Required But Not Provided

MAGLUMI Reaction Module	REF: 21160285
MAGLUMI Starter 1+2	REF: 130299004M
MAGLUMI Wash Concentrate	REF: 130299005M
MAGLUMI Light Check	REF: 130299006M



#### Preparation of the Reagent Integral

Before the sealing is removed, gentle and careful horizontal shaking of the Reagent Integral is essential (avoid foam formation!) Remove the sealing and turn the small wheel of the magnetic microbeads compartment to and fro, until the colour of the suspension has changed into brown. Place the Integral into the reagent area and let it stand there for 30 mins. During this time, the magnetic microbeads are automatically agitated and completely resuspended.

**Do not interchange integral component from different reagents or lot!**

#### Storage and Stability

- Sealed: Stored at 2-8°C until the expiry date.
- Opened: Stable for 4 weeks. To ensure best kit performance, it is recommended to place opened kits in the refrigerator if it's not going to be used on board during next 12 hours.



- Keep upright for storage.



- Keep away from direct sunlight.

## CALIBRATION AND TRACEABILITY

### 1) Traceability

To perform an accurate calibration, we have provided the test calibrators standardized against the SNIBE internal reference substance.

Calibrators in the reagent kit come from Sigma.

### 2) 2 Point Recalibration

Via the measurement of calibrators, the predefined master curve is adjusted (recalibrated) to a new, instrument-specific measurement level with each calibration.

### 3) Frequency of Recalibration

- After each exchange of lot (Reagent Integral or Starter Reagents).
- Every week and/or each time a new Integral is used

(recommendation).

- After each servicing of the MAGLUMI Fully Auto analyzer.
- If controls are beyond the expected range.

## SPECIMEN COLLECTION AND PREPARATION

Sample material: **Whole blood**

- Collect 5ml of venous blood in a glass tube without additives. Add 50µl 0.30mol/L EDTA to each tube. If you collect 2ml of venous blood in a tube, then add 20µl 0.30 mol/L EDTA to the tube, following the ratio.
- Overturn the tube up and down for several times, then place in room temperature then add the 0.83% NH<sub>4</sub>Cl to the plasma by 1:1 ratio. (200 µl plasma can be added 200 µl 0.83% NH<sub>4</sub>Cl) after it is mixed thoroughly for at least 30s.
- The whole blood sample are stable for up to 24 hours at 2-8°C. For longer storage, aliquot and store at -20°C (up to 30 days). Avoid repeated freezing and thawing.

#### Vacuum Tubes

- (a) Blank tubes are recommended type for collecting samples.  
 (b) Please ask SNIBE for advice if special additive must be used in sample collecting.

#### Specimen Conditions

- Do not use specimens with the following conditions:
  - heat-inactivated specimens;
  - Cadaver specimens or body fluids other than human serum;
  - Obvious microbial contamination.
- Use caution when handling patient specimens to prevent cross contamination. Use of disposable pipettes or pipette tips is recommended.
- Inspect all samples for bubbles. Remove bubbles with an applicator stick prior to analysis. Use a new applicator stick for each sample to prevent cross contamination.
- Serum and plasma specimens should be free of fibrin, red blood cells or other particulate matter.
- Ensure that complete clot formation in serum specimens has taken place prior to centrifugation. Some specimens, especially those from patients receiving anticoagulant or thrombolytic therapy, may exhibit increased clotting time. If the specimen is centrifuged before a complete clot forms, the presence of fibrin may cause erroneous results.

#### Preparation for Analysis

- Patient specimens with a cloudy or turbid appearance must be centrifuged prior to testing. Following centrifugation, avoid the lipid layer (if present) when pipetting the specimen into a sample cup or secondary tube.
- Specimens must be mixed THOROUGHLY after thawing by LOW speed vortexing or by gently inverting, and centrifuged prior to use to remove red blood cells or particulate matter to ensure consistency in the results. Multiple freeze-thaw cycles of specimens should be avoided.
- All samples (patient specimens, controls, and calibrators) should be tested within 3 hours of being placed on board the MAGLUMI System. Refer to the SNIBE service, for a more detailed discussion of onboard sample storage constraints.

#### Storage

- If testing will be delayed for more than 8 hours, remove serum or plasma from the serum or plasma separator, red blood cells or clot. Specimens removed from the separator gel, cells or clot may be stored up to 24 hours at 2-8°C.
- Specimens can be stored up to 30 days frozen at -20°C or colder.

#### Shipping

- Before shipping specimens, it is recommended that specimens be removed from the serum or plasma separator, red blood cells or clot. When shipped, specimens must be packaged and labeled in compliance with applicable state, federal and international regulations covering the transport of clinical specimens and infectious substances. Specimens must be shipped frozen (dry ice). Do not exceed the storage time limitations identified in this section of the package insert.

## WARNING AND PRECAUTIONS FOR USERS



- For use in *IN-VITRO* diagnostic procedures only.
- Package insert instructions must be carefully followed. Reliability of assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

### Safety Precautions

**CAUTION:** This product requires the handling of human specimens.

- The calibrators in this kit are prepared from bovine serum products. However, because no test method can offer complete assurance that HIV, Hepatitis B Virus or other infectious agents are absent; these reagents should be considered a potential biohazard and handled with the same precautions as applied to any serum or plasma specimen.
- All samples, biological reagents and materials used in the assay must be considered potentially able to transmit infectious agents. They should therefore be disposed of in accordance with the prevailing regulations and guidelines of the agencies holding jurisdiction over the laboratory, and the regulations of each country. Disposable materials must be incinerated; liquid waste must be decontaminated with sodium hypochlorite at a final concentration of 5% for at least half an hour. Any materials to be reused must be autoclaved using an overkill approach (USP 24,2000,p.2143). A minimum of one hour at 121°C is usually considered adequate, though the users must check the effectiveness of their decontamination cycle by initially validating it and routinely using biological indicators.
- It is recommended that all human sourced materials be considered potentially infectious and handled in accordance with the OSHA Standard on Bloodborne Pathogens<sup>13</sup>. Biosafety Level 214 or other appropriate biosafety practices should be used for materials that contain or are suspected of containing infectious agents.
- This product contains Sodium Azide; This material and its container must be disposed of in a safe way.
- Safety data sheets are available on request.

### Handling Precautions

- Do not use reagent kits beyond the expiration date.
- Do not mix reagents from different reagent kits.
- Prior to loading the Reagent Kit on the system for the first time, the microbeads requires mixing to re-suspend microbeads that have settled during shipment.
- For microbeads mixing instructions, refer to the KIT COMPONENTS, Preparation of the Reagent Integral section of this package insert.
- To avoid contamination, wear clean gloves when operating with a reagent kit and sample.
- Over time, residual liquids may dry on the kit surface, please pay attention the silicon film still exists on the surface of the kit.
- For a detailed discussion of handling precautions during system operation, refer to the SNIBE service information.

## TEST PROCEDURE

To ensure proper test performance, strictly adhere to the operating instructions of the MAGLUMI Fully Auto analyzer. Each test

parameter is identified via a RFID tag on the Reagent Integral. For further information please refer to the MAGLUMI Chemiluminescence Analyzer Operating Instructions.

80µl	Sample, calibrator or controls
+80µl	ABEI label
+20µl	Nano magnetic microbeads
15 min	Incubation
400µl	Cycle washing
3 s	Measurement

## DILUTION

Sample dilution by analyzer is not available in this reagent kit. Samples with concentrations above the measuring range can be diluted manually. After manual dilution, multiply the result by the dilution factor.

Please choose applicable diluents or ask SNIBE for advice before manual dilution must be processed.

## QUALITY CONTROL

- Observe quality control guidelines for medical laboratories
- Use suitable controls for in-house quality control. Controls should be run at least once every 24 hours when the test is in use, once per reagent kit and after every calibration. The control intervals should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined ranges. Each laboratory should establish guidelines for corrective measures to be taken if values fall outside the range.

## LIMITATIONS OF THE PROCEDURE

### 1) Limitations

Patients with malignancies may exhibit Cyclosporine A values within the normal range. Cyclosporine A concentrations may be elevated in case of liver cirrhosis, hepatitis or tyrosinaemia. Thus, Cyclosporine A determination is more suitable for therapeutic monitoring and follow-up as well as for a comparison with histological results. Cyclosporine A serum levels may only be interpreted in context with the clinical picture and other diagnostic procedures. The Cyclosporine A assay should not be used as the only criterion for cancer screening.

### 2) Interfering Substances

No interference with test results is seen by concentrations of bilirubin <0.06mg/ml, haemoglobin <16mg/dl or triglycerides <12.5mg/ml.

### 3) HAMA

Patient samples containing human anti-mouse antibodies (HAMA) may give falsely elevated or decreased values. Although HAMA-neutralising agents are added, extremely high HAMA serum concentrations may occasionally influence results.

## RESULTS

### 1) Calculation of Results

The analyzer automatically calculates the CSA concentration in each sample by means of a calibration curve which is generated by a 2-point calibration master curve procedure. The results are expressed in ng/ml. For further information please refer to the MAGLUMI Chemiluminescence Analyzer Operating Instructions.

### 2) Interpretation of Results

- Results of study in clinical centers with group of individuals, 95% of the results were:100-400ng/ml.
- Results may differ between laboratories due to variations in population and test method. If necessary, each laboratory should establish its own reference range.

## PERFORMANCE CHARACTERISTICS

### 1) Precision

Intra-assay coefficient of variation was evaluated on 3 different levels of control serum repeatedly measured 20 times in the same run, calculating the coefficient of variation.

<b>Intra-assay precision</b>			
Control	Mean(ng/ml)	SD(ng/ml)	CV%
Level 1	120.33	6.53	5.43
Level 2	357.96	20.48	5.72
Level 3	1239.81	67.94	5.48

Inter-assay coefficient of variation was evaluated on three batches of kit. Repeatedly measured 3 different levels of control serum 21 times, calculating the coefficient of variation.

<b>Inter-assay precision</b>			
Control	Mean(ng/ml)	SD(ng/ml)	CV%
Level 1	130.57	12.76	9.77
Level 2	347.24	32.71	9.42
Level 3	1249.73	114.23	9.14

## 2) Analytical Sensitivity

The sensitivity is defined as the concentration of CSA equivalent to the mean RLU of 20 replicates of the zero standard plus two standard deviations corresponding to the concentration from the standard curve. The sensitivity is typically less than 12.5ng/ml.

## 3) Specificity

The specificity of the CSA assay system was assessed by measuring the apparent response of the assay to various potentially cross reactive analytes.

Compound	Concentration	Cross reactivity
AM1(M17)	500ng/ml	1.6%
AMIC(M18)	250ng/ml	1.3%
AM9(M1)	250ng/ml	1.1%
AM19(M8)	250ng/ml	1.5%

## 4) Recovery

Consider calibrator high of known concentration as a sample, dilute it by 1:2 ratio with diluents, and measure its diluted concentration for 10 times. Then calculate the recovery of measured concentration and expected concentration. The recovery should be within 90% -110%.

Expected	Mean Measuring	Recovery
630.54ng/ml	614.23ng/ml	97%

## 5) Linearity

Use CSA calibrator to prepare the six point standard curve, measuring all points' RLU except point A, and then do four parameter linear fitting in double logarithm coordinate, the absolute linear correlation coefficient(r) should be bigger than 0.9800.

Calibrator Point	Concentration ng/ml	Absolute linear correlation coefficient (r)
A	0	
B	50	r=0.9841
C	200	
D	500	
E	1000	
F	2000	

## REFERENCES

- [Chinese Journal of Hospital Pharmacy] Jiaozheng, Zhangming, Zhonglongjin, Shixiaojin, Zhongmingkang
- Study on the correlation of results between AxSYM and TDx monoclonal whole blood cyclosporine assay [ Chinese Journal of Hospital Pharmacy] Yuzicheng, Shixinmin, Dongyongqin, Wangxianghui, Xuda.
- Comparison of high-performance liquid chromatography and fluorescence polarization immunoassay in measuring of whole blood cyclosporin A [Jilin Medical Journal], ZHAO Hong , RONG Mo-ke , LI Ya-ping , LANG Zhi-gang